

BID EVALUATION REPORT FOR THE PURCHASE OF LAB KITS

Name of Item	Allocation of funds	Date of advertisement of tender or letter inviting bids from bidder	No. of Tender Document Sold	No. of tender submitted	Closing date of tender	Opening date of tender	Details/No. of bidder participated in bidding	Name of the qualified bidder and disqualification		Reasons for disqualification	Financial bid opening date	Seniority according to evaluation
								Qualified	Disqualified			
1	2	3	4	5	6	7	8	9	10	11	12	13
Purchase of PCR & Extraction Kit and Consumables.	Rs.19,526,045/-	11/7/2021	6	6	2/8/2021	2/8/2021	1-M/S Hooraa Pharma 2-M/S Ayyan Molecular 3-M/S Global Marketing 4- M/S Popular International 5- M/S Pakistan Hospital 6-M/S Pakistan Microbiological Associates	1-M/S Ayyan Molecular 2-M/S Global Marketing 3- M/S Popular International 4- M/S Pakistan Hospital	1-M/S Hooraa Pharma 2-M/S Pakistan Microbiological Associates	(1) M/S Hooraa Pharma (Kits quoted by the firm (Sansure) did not give satisfactory results and Extraction kits not compatible for ten beads extraction system). (2) M/S Pakistan Microbiological (The quoted kits of the said firm (Systaaq USA) not FDA Approved.	29-11-2021	1-M/S Ayyan Molecular 2-M/S Global Marketing 3- M/S Popular International 4- M/S Pakistan Hospital
Kits & Reagents for Immunoassay Special Chemistry (With Machine on Reagent Rental Basis and all consumables item, Controls, Calibrators, Services)	Rs.15,606,734/-	11/7/2021	4	4	2/8/2021	2/8/2021	1-M/S Diagnostic Medical 2-M/S Sind Medical 3-M/S Pakistan Microbiology 4-M/S Hooraa Pharma	1-M/S Sind Medical 2-M/S Pakistan Microbiology 3-M/S Hooraa Pharma	1-M/S Diagnostic Medical	1-- Sample Volume 10ul to 80ul instead of 150ul to 250ul - upto 20 Integrated Reagent on-board Instead of upto 24 Reagents on-board at a time. The Said firm did not submitted their technical bid with proper book binding & proper page marking. Therefore their bid not considered as per tender & Bidding documents clauses.	29-11-2021	1- M/S Hooraa Pharma 2-M/S Pakistan Microbiology 3-M/S Sind Medical Store
Kits & Reagents for Coagulation Analyzer (With Machine on Reagent Rental Basis and all consumables item, Controls, Calibrators, Services) for Central Lab.	Rs.25,674,940/-	11/7/2021	5	5	3/8/2021	3/8/2021	1- M/S S.Ejaz Ud Din 2- M/S A.M Sales 3- M/S Hasnain Scientific 4- M/S popular International 5- M/S Biomedics	1- M/S S.Ejaz Ud Din 2-M/S AM Sales	1-M/S Hasnain Scientific 2- M/S popular International 3- M/S Biomedics	1- Demonstration done in Pvt Setup. Working Very low not compatible to heavy workload of LGH so performance cannot be assessed. Lahore Based Biomedical Engineer not available, frequent troubleshooting is required due to heavy workload, necessitated 24/7 availability of biomedical engineer. Previous Performance not satisfactory in LGH. Technically not approved. 2- Demonstration done in Jinnah Hospital. Two Machines installed by Company to cover workload. QC data not available in Machine. Provision of regular controls not documented. Demonstration not satisfactory. 3- Letter was issued to the firm vide No.64353-57/LGH dated 16-10-2021, but the firm failed to demonstrate the quoted item.	29-11-2021	1- M/S S.Ejaz Ud Din 2-M/S AM Sales

Signature
General Hospital
 Purchase

Signature

Name of Item	Allocation of funds	Date of advertisement of tender or letter inviting bids from bidder	No. of Tender Document Sold	No. of tender submitted	Closing date of tender	Opening date of tender	Details/No. of bidder participated in bidding	Name of the qualified bidder and disqualification		Reasons for disqualification	Financial bid opening date	Seniority according to evaluation
								Qualified	Disqualified			
Kits & Reagents for Coagulation Analyzer (With Machine on Reagent Rental Basis and all consumables item, Controls, Calibrators, Services) for Emergency Lab.	Rs.6,268,800/-	11/7/2021	6	6	3/8/2021	3/8/2021	1- M/S S.Ejaz Ud Din 2- M/S A.M Sales 3- M/S Hasnain Scientific 4- M/S popular International 5- M/S Biomedics 6- M/S Diagnostic Medical	1- M/S S.Ejaz Ud Din 4- M/S A.M Sales	1- M/S Hasnain Scientific 2- M/S popular International 3- M/S Biomedics 4- M/S Diagnostic Medical	1- Demonstration done in Pvt Setup. Working Very low not compatible to heavy workload of LGH so performance cannot be assessor. Lahore Based Biomedical Engineer not available, frequent troubleshooting is required due to heavy workload, necessitated 24/7 availability of biomedical engineer. Previous Performance not satisfactory in LGH. Technically not approved. 2- Demonstration done in Jinnah Hospital. Two Machines installed by Company to cover workload. QC data not available in Machine. Provision of regular controls not documented. Demonstration not satisfactory. 3- Letter was issued to the firm vide No.64353-57/LGH dated 16-10-2021, but the firm failed to demonstrate the quoted item. 4- 3- Letter was issued to the firm vide No.64353-57/LGH dated 16-10-2021, but the firm failed to demonstrate the quoted item.	29-11-2021	1- M/S S.Ejaz Ud Din 2- M/S AM Sales
Purchase of Routine Chemistry Kits & Reagents (With Machine on Reagent Rental Basis and all consumables item, Controls, Calibrators, UPS, RO Plant, Services) for Emergency Lab.	Rs.7,796,500/-	21-8-2021	4	4	14-9-2021	14-9-2021	1- M/S Biocept International 2- M/S Popular International 3- Pakistan Microbiological 4-M/S SKI Technologies	1- Pakistan Microbiological	1- M/S Biocept International 2- M/S Popular International 3- M/S SKI Technologies	1- Technical Assessment Criteria not fulfill:- - Acceptance if terms and condition, tender documents duly signed and stamped, not signed. - Technical Bid not attached. - Valid Professional Tax Certificate not attached. - Photocopy of 2% CDR/Bank Guarantee/Pay Order/demand draft not defaced. 2- Technical Assessment Criteria not fulfill:- - Relevant Experience of Quoted item not attached. Following Deficiencies in technical specification:- - Sample type: Serum, Plasma, Body Fluids etc Not Verified from Broucher. - Automatic Sample Quality Analysis for Lipemia, Hemolysis Icterus not verified from Broucher. - 290 test/hr instead of 300 test/hr - Temperature Control for assays at 16-30c instead of 37c. - Automatic flagging of results outside user defined limits, Not verified from Broucher. - Compatible Imported Online Sine wave UPS with Battery backup for 30 minutes(Emerson, Liebert, Chloride, MGE, APC or Equivalent) not offer. 3- Technical Assessment Criteria not fulfill:- - Relevant Experience of Quoted item not attached. Following Deficiencies in technical specification:- - Body Fluid, Not verified from Broucher. - Automatic Sample Quality Analysis for Lipemia, Hemolysis Icterus not verified from Broucher. - Sample Volume 2-50ul not Verified from Broucher. - Temperature Control for assays at 15-32 instead of 37c.	4/12/2021	1- M/S Pakistan Microbiological Associates

Signature
TMS Purchase
Lahore General Hospital

TH