

Minutes of Technical Evaluation Committee Meeting

Held 19-4-2024

1. The Technical Evaluation Committee meeting was held on 19-4-2024 under the chairmanship of Prof. Dr. Muhammad Shahid, Prof of Peads Lahore General Hospital, Lahore, regarding technical evaluation of the offers for the Purchase of Coagulation Kits, HbA1c kits with Machines on reagent rental basis and cost of other store items of this hospital.

2. Following committee members attended the meeting:

Prof. Dr. Akif Dilshad, Prof of Gastroenterology

Dr. M. Shabbir Ahmad, AP Surgery

Dr. Khurram Butt, AMS (Stores)

Dr. Zeeshan Hadder, AMS (Purchase)

Dr. Irem Iqbal, AP Pathology

Dr. M. Riaz, AP Pathology

Dr. Kashif Mughal, APMO Pathology

Dr. Saira Akhlaq, DMS (Purchase)


Eng. Ali Natiq Zaidi, Bio Medical Engineer

3. The technical sheets were prepared by the end-user & Technical Officer, Health Department, Lahore and was placed before the committee.

4. The committee decided that the clauses of tender enquiry & PPRA rules 2014, will be followed strictly.

5. The following items were discussed:-

Sr #	Name of Item	Name of Firm	Remarks	Decision
1	Coagulation Kits with Machine on Reagent rental basis	M/S S.Ejaz Ud Din	(Model:-CS-1600 Sysmex Corporation Japan) As Per Bidding Documents & tender Specification. Demonstration Satisfactory.	Responsive
		M/S SKI Technologies	(Model:-Star Max Compact 3 Stago france) As Per Bidding Documents & tender Specification. Demonstration Satisfactory.	Responsive
		M/S Popular International	Technical Evaluation Criteria Not Fullfill:- - (Model;- Humaclot PRO Human Germany) - The Quoted items must bear FDA510K. CE (MDD) or MHLW (Ministry of Health, Labor and welfare). Two Certificate are mandatory for reagent rental tenders. The said firm have only CE certificate, not dual certification. - Speed S/M, not Compatible with LGH work load. QC/Control data not available in machine.	Non Responsive



Sr #	Name of Item	Name of Firm	Remarks	Decision
2	HbA1C kits with machine on reagent rental basis	M/S Z.A Biomedics	Model:- G8 Tosoh Japan Both bidders participated with same and model, which is violation of PPRA rules.	Non Responsive
		M/S PK Medi Engineering	- Tosho issued authorization to M/S Z.A Biomedics. - Tosho Issued authorization to M/S Hitec and Hitech authorized to M/S PK Medi Engineering.	
		M/S Chemical House	As per bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Ayyan Molecular	Model:- GSH-60 Goldsite Diagnostic Japan/China The said firm did not demonstrate the quoted model as per requirement of the End user and tender documents.	Non Responsive
		M/S Sind Medical Store	Technical Specification Not Fullfill:- (Model;- Premier Hb9210TM Trinity Biotech Ireland) - Throughput 48test/hr instead of 60 test/hr or more	Non Responsive
		M/S Sure Bio Diagnostic	Technical Evaluation Criteria Not Fullfill:- - (Model;- GH900 Plus Lifetronic China) - The Quoted items must bear FDA510K. CE (MDD) or MHLW (Ministry of Health, Labor and welfare). Two Certificate are mandatory for reagent rental tenders. The said firm have only one certificare, not dual certification. - Throughput 29 test/hr instead of 60test/hr or more	Non Responsive
3	Cost of Other Store Items	M/S Imran & CO	The Sample of the quoted item " Methyleated Spirit" not approved.	Non Responsive
		M/S Z.A Biomedics	The Sample of the quoted item " Methyleated Spirit" not approved.	Non Responsive
		M/S Vertxe Medical	The Sample of the quoted item "Soda Lime 05kg" approved.	Responsive
		M/S Ever Clean	The Sample of the Quoted items"Methyleated Spirit" approved.	Responsive
		M/S Omer Health	The Sample of the quoted item "Soda Lime 05kg" approved.	Responsive
		M/S Skardu Surgical	The Sample of the quoted items "Sterlization Strip-V Vapor Control Italy" Approved.	Responsive
		M/S Fab Healthcare	The Sample of the quoted items "Sterlization Strip-V Astra Med Uk" Approved and the sample of Ambo Bag not provided by the firm for sample evaluation.	Responsive for sterilization strip.

Sr #	Name of Item	Name of Firm	Remarks	Decision
------	--------------	--------------	---------	----------

The working paper for Technical Evaluation is prepared by the DMS (Purchase) Dr.Saira Akhlaq and Technical Officer BME Engr. Ali, Natiq Zaidi, on the basis of hospital Bidding documents & documents Provided by bidder. We Will responsible if any discrepancy at any stage later.

for
Dr. Saira Akhlaq DMS (Purchase)



Mr. Ali Natiq Zaidi, Biomedical Engineer



Prof. Dr. M. Shahid, Prof of Peads, Chairman Committee




Prof. Dr. Akif Dilshad, Prof of Gastroenterology




Dr. M. Shabbir Ahmad, AP Surgery




Dr. Khurram Butt, AMS (Stores)



Dr. Zeeshan Hadder, AMS (Purchase)



Dr. Irem Iqbal, AP Pathology




Dr. M. Riaz, AP Pathology



Dr. Kashif Mughal, APMO Pathology



for
Dr. Saira Akhlaq, DMS (Purchase)



Eng. Ali Natiq Zaidi, Bio Medical Engineer

