


**Minutes of Technical evaluation Committee Meeting
Held 10-12-2024**

1. The Technical Evaluation Committee meeting was held on 10-12-2024 under the chairmanship of Prof. Dr. Muhammad Shahid, Prof of Peads Lahore General Hospital, Lahore, regarding technical evaluation of the offers for the Purchase of Routine Chemistry Kits & Special Chemistry Kits with Machine on reagent rental basis for this hospital.
2. Following committee members attended the meeting:
 - Prof. Dr. Akif Dilshad, Prof of Gastroenterology
 - Prof. Dr. Nazish Saqlain, Prof of Pathology
 - Dr. Roomana Akhlaque, Associate Prof of Plastic Surgery
 - Dr. M. Shabbir Ahmad, AP Surgery
 - Dr. M. Riaz, AP Pathology
 - Dr. Zeeshan Hadder, AMS (Purchase)
 - Engr. Ali Natiq Zaidi, Bio Medical Engineer
3. The technical sheets were prepared by the end-user & Technical Officer, Health Department, Lahore and was placed before the committee.
4. The committee decided that the clauses of tender enquiry & PPRA rules 2014, will be followed strictly.
5. The following items were discussed:-

Sr #	Name of Item	Name of Firm	Remarks	Decision
1	Special Chemistry Kits with Machine on reagent rental basis	M/S ZEDCO	The said firm did not quote the complete Package of Special Chemistry Kits as per bidding documents. "Growth Hormone" not quoted.	Non Responsive
		M/S Nextek Healthcare	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Pakistan Microbiological Associates	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Roche Pakistan	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Sind Medical Store	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Hooraa Pharma	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Pakistan Microbiological Associates	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Roche Pakistan	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive



2	Routine Chemistry Kits with Machine on reagent rental basis.	M/S Hoora Pharma	As per Bidding Documents & Tender Specification. Demonstration Satisfactory.	Responsive
		M/S Sind Medical Store	Technical Evaluation Criteria Not fulfill:- - The quoted items must bear FDA510K, CE (MDD), or MHLW (Ministry of Health, Labor and welfare) (Two certificates are mandatory for reagent rental tenders). "The firm attached only 01 certificate instead of Dual Certificate" - The offer of the said firm is ambiguous (Analyzer of SEMS Lab not mentioned, Moreover they have quoted backup for Central Lab Model BS-430. Analyzer for SEMS Lab not quoted, nor its certification & technical data attached.	Non Responsive
		M/S Popular International	Technical Evaluation Criteria Not fulfill:- - The quoted items must bear FDA510K, CE (MDD), or MHLW (Ministry of Health, Labor and welfare) (Two certificates are mandatory for reagent rental tenders). "The firm attached only 01 certificate instead of Dual Certificate". - The Quoted Model for Central Lab "PICTUS-770" did not fulfill the required throughput i.e."800test/hr".	Non Responsive

Prof. Dr. M. Shahid, Prof of Peads, Chairman Committee



Prof. Dr. Akif Dilshad, Prof of Gastroenterology


Prof. Dr. Nazish Saqlain, Prof of Pathology



Dr. Roomana Akhlaque, Associate Prof of Plastic Surgery



Dr. M. Shabbir Ahmad, AP Surgery



Dr. M. Riaz, AP Pathology



Dr. Zeeshan Hadder, AMS (Purchase)



Engr. Ali Natiq Zaidi, Bio Medical Engineer

