



OFFICE OF THE MEDICAL SUPERINTENDENT
LAHORE GENERAL HOSPITAL, LAHORE

PH: 042-99268836, 99268842, Exch: 99268801-8

No. 16075 /LGH, Lahore Dated 01-03 /2022

To


The Secretary,
Govt. of the Punjab,
SHE & ME Department,
Lahore

Attention: Mr. Abdul Wahab
(Focal Person) PPRA Website,
Health Department, Govt. of the Punjab,
Lahore.

Subject: **MINUTES OF MEETING.**

With reference to the subject cited above.

It is requested to issue an instruction to the concern focal person for uploading the minutes of Grievance Committee meeting held on 01-03-2022 on official website of Health Department regarding purchase of Diathermy Machine Dedicated for Endoscopy System through ADP Scheme "Replacement of Equipment of Gastroenterology, Lahore General Hospital, Lahore".



Medical Superintendent
Lahore General Hospital
Lahore

No. _____ /LGH

Dated _____ 2022

A copy is forwarded for information to the:-

1. Principal, PGMI/AMC/ Lahore General Hospital, Lahore.
2. Prof. Dr. Muhammad Moin, Prof of Ophthalmology, PGMI/AMC/LGH, Lahore.
3. Prof. Dr. Ghias-un-Nabi Tayyab, Prof of Medicine PGMI/AMC/LGH, Lahore.
4. AMS/DMS (ADP Purchase), LGH, Lahore.
5. Computer Programmer (with the request to upload the minutes on official website of LGH, Lahore.


Medical Superintendent
Lahore General Hospital
Lahore

LAHORE GENERAL HOSPITAL, LAHORE

1. Grievance committee meetings were held on following dates under the chairmanship of Prof. Dr. Muhammad Moin (Prof. of Ophthalmology), AMC/PGMI/LGH, Lahore to resolve the grievances submitted by the firm, for the purchase of ADP Scheme "Replacement of Equipment of Gastroenterology, Lahore General Hospital, Lahore".
2. Following committee members attended the meeting:
Prof. Dr. Farah Shafi, Prof. of Medicine, AMC/PGMI/LGH, Lahore.
Prof. Dr. Muhammad Shahid, Prof. of Peads, AMC/PGMI/LGH, Lahore.
Dr. Arif Shahzad Bhatti, Private Member
Mr. Muhammad Irfan, Biomedical Engineer, PGM/AMC, LHR
3. The Representatives of the firms are also attend the meeting
4. The items were discussed in detail as follows:

SR #	NAME OF ITEM	NAME OF FIRM	GRIEVANCE OF THE M/S ENDO-KARE	GRIEVANCE COMMITTEE 22-02-2022	GRIEVANCE COMMITTEE 01-03-2022	DECISION
1	Diathermy Machine	M/S Endo Kare	Respected sir, We hereby submitted our reservation against the technical scrutiny of the tender against <u>M/S Popular International</u> ➤ The quoted model is WEM SS-501SX is manufactured by WEM Equipmentos Electronicos Ltda in Brazil therefore WEM country of manufacturer and country of origin both are Brazil. ➤ In this tender a dedicated unit for	➤ The Representative of M/S Endo Kare and M/S Popular International attended the meeting and asked to M/S Endo Kare to explain their Grievance. ➤ He explained that the quoted model WEM SS-501SX is manufacturer and country of origin is Brazil. Furthermore the quoted model does not fulfill the tender specification and quoted model doesn't have GASTROCU/ENDOCUT / PULSECU functions, so requested to reject the bid/tender of the said firm.	➤ The Committee called the Representative of M/S Popular International and asked him that you were directed to give demonstration of the Quoted Model by the End-User & Performance Report of Gastro Department of any Institution within the Province of the Punjab and FDA 510K of quoted model till 24-02-2022. In this regard, a letter vide NO.15028/LGH dated 25-02-2022 was issued to your Firm to provide the above requisites before 28-02-2022. ➤ The Representative of the Firm informed to the	Grievance of M/S Endo-Kare accepted and M/S Popular International Technically Rejected

<p>endoscopy was required, whereas the quoted model has 400 watts output designed for surgical applications not suitable for endoscopic applications. There is a great risk of perforation in the small intestine if the wattage were not set to lower level i.e. less than 120 w.</p> <p>➤ Moreover, the heat generated by 400w electrosurgical generator may burn the endoscope distal end that can damage the CCD chip-on-tip and such damage by third party diathermy units/end-therapy device is not covered in warranties.</p> <p>➤ The quoted unit does not have the GASTRO/CUT/END OCUT/PULSECUT functions as required in specifications.</p>		
<p>➤ The Representative of M/S Popular International explain their case, he explained that WEM Equipamentos Electronicos Ltda Brazil is a part of - Medtronic plc group of companies. Furthermore quoted model 100% meets the tender specification.</p> <p>➤ The committee called the Dr. Ghias-ul-Hassan, Medical Unit-1 PGM/AMC/LGH, LHR for comments. (He attended the meeting on the behalf of Professor Ghias-un-Nabi Tayyab Head of Medical Unit-1 PGM/AMC/LGH, LHR)</p> <p>➤ After detailed discussion he requested to the committee to ask the firm to arrange the demonstration of the quoted model within two days i.e. 24-02-2022 for confirmation working & safety in endoscopic procedure.</p> <p>➤ The Technical Officer/ Biomedical Engineer PGM/AMC/LGH, LHR asked the Chairman to</p>	<p>Committee that</p> <ol style="list-style-type: none"> 1. Quoted Model were demonstrated to the End User. 2. FDA 510K is not mandatory, our manufacturer WEM Equipamentos Electronicos Ltda Brazil is a part of - Medtronic plc group of companies Ltd. did not sale the said item in USA however required CE is attached in the Bid. <p>➤ After detailed discussion and going through the case the Committee found that M/S Popular International did not provide the performance reports of Quoted Model from Gastro Department of any Institution within the Province of the Punjab before 28-02-2022. Hence it was decided said Firm is declared Technically Non Responsive.</p>	

W

Arshad

			<p>issue the instructions to the firm to provide the two performance reports of the Gastro Department of any institution within the province of Punjab and certificate of FDA 510k of the quoted model before the above mentioned date.</p> <p>➤ The Committee directed the representative of the firm (M/S Popular International) to arrange the demonstration of the quoted Model and also directed to provide two performance reports of the quoted model from Gastro Department of any institute within the province of Punjab and certificate of 510k before the 24-02-2022. The representative of M/S Popular International agreed that he will provide the requisites before above mentioned date.</p>		
--	--	--	---	--	--

Prof. Dr. Muhammad Moin Prof. of Ophthalmology, AMC/PGMILGH, Lahore

Prof. Dr. Farah Shafi, Prof. of Medicine, AMC/PGMILGH, Lahore.

Wain
Farah

Prof. Dr. Muhammad Shahid, Prof. of Peds, AMC/PGMI/LGH, Lahore.

Dr. Arif Shahzad Bhatti, Private Member

Mr. Muhammad Irfan, Biomedical Engineer, PGMI/AMC, LHR

